

MICROBIOLOGICAL PERFORMANCE QUALIFICATION OF THE ROBOTIC SYSTEMS APOTECA-SYRINGE AND APOTECA-UNIT

Background and Importance

Fully automated robotic systems designed for the aseptic preparation of non-cytotoxic ready-to-administer (RTA) and ready-to-use (RTU) parenterals require microbiological performance qualification during the implementation process.

Aim and Objectives

The objective of the study was the microbiological performance qualification of the fully automated robotic systems APOTECAsyringe and APOTECAunit (Loccioni, Italy) by media-fill simulations and supplemental environmental monitoring in the critical zones, followed by a cleaning procedure and 4 hour-UV radiation regarding the user manual.

Materials and Methods

APOTECAsyringe – Media-fill simulation; automated filled, capped, labelled

- Product: Polypropylene syringe 10 mL,
n = 100/day x 5 days (total: 500 syringes)**

Content: premixed media: 1500 mL Tryptic Soy Broth Single Strength
750 mL BD Tryptic Soy Broth Double Strength + 750 mL Aqua ad inject.

2. Environmental monitoring




Figure 1: Type and location of environmental monitoring

C1-3=contact plates, S1-3=settle plates,
P=particle counting ($\geq 0.5 \mu\text{m}/\text{m}^3$, $5 \mu\text{m}/\text{m}^3$)

APOTECAunit – Media-fill simulation; automated filled, capped, labelled

- Product: Polypropylene syringe 50 mL, n = 25/day x 10 days (total: 250 syr.)**
Content: 25 mL BD Tryptic Soy Broth Double Strength + 25 mL Aqua ad inject.

- Product: Polyolefine bag 100 mL, n = 25/day x 10 days (total 250 bags)**

Content: 50 mL prefilled 0.9% NaCl infusion bag freeflex® Fresenius
+ 50 mL BD Tryptic Soy Broth Double Strength

3. Environmental monitoring




Figure 2: Type and location of environmental monitoring

C1-8=contact plates, S1-4=settle plates, A1-3=active air sampling,
P1-2=particle counting ($\geq 0.5 \mu\text{m}/\text{m}^3$, $5 \mu\text{m}/\text{m}^3$),

- Incubation of media-fills:** At 15°C - 25°C **Analysis:** Visual inspection of turbidity; Day 7, Day 14
- Environmental controls:** Particle counting (PZG 131); active air sampling (MAS-100 NT) and settle plates (CASO-Agar m. LT-ICR 30 mL IPC, heipha/Merck); contact plates and fingerprints (CASO-Abklatschagar Tryptic Soy Contact Agar LT-ICR, heipha/Merck)
Sampling locations see figure 1 and figure 2
 - Incubation of plates:** 5 - 7 days at 15°C - 25°C followed by 2 - 3 days at 30°C - 35°C
 - Analysis:** Colony forming units (CFU) counting, particle counting; acceptance criteria: EU-GMP Guide, Annex 1 (2007)

Results

APOTECAsyringe

Table 1: Results of visual inspection of media-fill products and supplemental environmental monitoring
CFU=colony forming units, FR=fingerprints right hand, FL=fingerprints left hand.

Media-fill products	Environmental Monitoring - type and location											
	Turbidity		Fingerprints [CFU/plate]		Contact plates [CFU/plate]			Settle plates [CFU/plate]			Particles [$\mu\text{m}/\text{m}^3$]	
	Day 7	Day 14	FR	FL	C1	C2	C3	S1	S2	S3	$\geq 0,5$	≥ 5
Day 1	None	None	0	0	0	0	2	0	0	1	69	1
Day 2	None	None	0	0	0	0	0	0	0	0	0	0
Day 3	None	None	0	0	0	0	0	0	0	0	3	1
Day 4	None	None	1	0	3	0	1	0	0	0	85	0
Day 5	None	None	0	0	0	0	0	0	0	0	30	1
Mean	-	-	0.2	0	0.6	0	0.6	0	0	0.2	37.4	0.6

APOTECAunit

Table 2: Results of visual inspection of media-fill products and supplemental environmental monitoring
CFU=colony forming units, FR=fingerprints right hand, FL=fingerprints left hand.

Media-fill products	Environmental Monitoring – type and location																						
	Turbidity		Fingerprints [CFU/plate]		Contact plates [CFU/plate]								Settle plates [CFU/plate]				Active air sampling [CFU/plate]			No. Particles [$\mu\text{m}/\text{m}^3$]			
	Day 7	Day 14	FR	FL	C1	C2	C3	C4	C5	C6	C7	C8	S1	S2	S3	S4	A1	A2	A3	$\geq 0,5$	≥ 5	$\geq 0,5$	≥ 5
Day 1	None	None	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	24	2	24	1
Day 2	None	None	3	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	10	12	5	17	0
Day 3	None	None	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	35	4	15	1
Day 4	None	None	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	74	3	0	0
Day 5	None	None	2	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	5	0	81	8
Day 6	None	None	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	22	2	10	0
Day 7	None	None	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	6	0	19	0
Day 8	None	None	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	3	0	29	1
Day 9	None	None	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	37	3	7	0
Day 10	None	None	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	2	0	7	2
Mean	-	-	0.6	0	0	0.1	0	0.1	0	0	0	0	0	0	0.2	0	0.1	1	22	1.9	20.9	1.3	

None of the 1.000 media-fill products showed turbidity when inspected after 7 and 14 days of incubation, thereby indicating no growth of microorganisms.

CFU counts and particle numbers met the acceptance criteria of EU-GMP Guide, Annex 1 (2007), cleanroom class A, except loading area (cleanroom class B).

Conclusion

The fully automated robotic systems APOTECAsyringe and APOTECAunit passed the microbiological performance qualification.

Both robotic systems allow safe automated aseptic preparation of non-cytotoxic RTA and RTU parenteral products.

